

ENG M62-

Chief, Operations Division, Medical Staff

Chief, Engineering Staff, OC

Status of Radiological Survey and Radiation Detection Instruments
Quarterly Report

1. During July, August, and September 1962, the radiological survey and radiation detection instruments [redacted] were tested to determine their operational status. A total of five instruments were found to be inoperative. These instruments were placed in operating condition.

2. The remainder of the tubes and batteries were received. The installation of the batteries is 90% complete.

3. The remainder of the battery installation and calibration of the instruments will commence as soon as time is available.

4. A total of thirteen man-hours has been expended during the last three months.

[redacted]
R&D/Lab/JWM/rkb (20 September 1962)

Distribution:

- Original and 1 - Addressee
- 1 - Lab Subj.
- 1 - OC-E Chrono
- 1 - Dev/s
- 1 - Lab Pending ←

